

CERTIFICATE of CONFORMANCE

PRODUCT INFORMATION	
Date of Certification:	2025-12-03
BPS Re-Order ID:	01GOGG2230RST
Product Description:	GOGGLES, BASIC 2230R,CLEAR, WITH RUBBER HEAD STRAP, 1/PACK, STERILE
Batch Control ID:	13077050
Expiration Date:	2028-11-02
We certify that this product was manufactured in accordance to all described specifications below, as indicated on the STERIS BPS PRODUCT SPECIFICATION (Ref BPS-PS022).	

Product Specification	Requirement	Test Method
Aesthetics – Threads	Free from detached threads on inner and outer surfaces	BPS-08-006-08 Visual Inspection
Aesthetics – Seams	Free from holes, splits, gaps or other signs of damage	BPS-08-006-08 Visual Inspection
Aesthetics – Foreign Particulate	Free from dirt, stains, and other foreign material (embedded or detached)	BPS-08-006-08 Visual Inspection
Aesthetics – Color	White (see Note 1)	BPS-08-006-08 Visual Inspection
Product Dimensions	Cut, sewn and/or sealed and/or packaged to conform to BPS Product Specification Sheet dimensions and tolerances	BPS-08-006-09 Dimensional Inspection

Note 1: Product color may vary due to natural variation in materials. Tyvek may vary from white to light pink or yellow based on natural oxidation of the material. STERIS Corp has verified that this phenomenon has no effect on the functionality of the product.

Signature:



Quality Representative
STERIS Corporation